

Infectious Disease Control IV – Level 3

Assignment of values (**LOT** 15C0701)

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 3	
			Target	Range
Auto Lumo A2000/Auto Lumo A2000 plus	Anti-HBs	mIU/mL	102.17	71.52-132.83
	Anti-HBe	PEI U/mL	0.98	0.69-1.28
	Anti-HBc	PEI U/mL	7.83	5.48-10.18
	HIV Ag/Ab combo	S/CO	6.32	4.42-8.21

Instrument	Analyte	Unit	Infectious Disease Control IV-Level 3	
			Target	Range
Auto Lumo A1000	Anti-HBs	mIU/mL	91.94	64.35-119.52
	Anti-HBe	PEI U/mL	1.08	0.76-1.41
	Anti-HBc	PEI U/mL	8.62	6.03-11.21
	HIV Ag/Ab combo	S/CO	5.91	4.14-7.69

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.