

Infectious Disease Control III– Level 1

Assignment of values (LOT 22K0701)

Instrument	Analyte	Unit	Infectious Disease Control III-Level 1
			Range
Auto Lumo A2000/Auto Lumo A2000 plus	HBsAg	IU/mL	<0.05
	HBeAg	IU/mL	<0.10
	Anti-TP	S/CO	<1.00
	Anti-HCV	S/CO	<1.00
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	<1.00
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	<1.00
Auto Lumo A1000	HBsAg	IU/mL	<0.05
	HBeAg	IU/mL	<0.10
	Anti-TP	S/CO	<1.00
	Anti-HCV	S/CO	<1.00
	Anti-HIV (Anti-HIV CLIA Microparticles)	S/CO	<1.00
	Anti-HIV (HIV Ag/Ab Combo CLIA Microparticles)	S/CO	<1.00

Remark: The range printed in the assigned value table should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.