

Cardiac Markers Control– Level 2

Assignment of values (**LOT** 2022010601)

Instrument	Analyte	Unit	Cardiac Markers Control-Level 2	
			Target	Range
Abbott I2000sr	BNP	pg/mL	312.63	218.84-415.80
	CK-MB	ng/mL	59.30	41.51-78.87
	cTnl	ng/mL	1.89	1.33-2.52
	HCY	μmol/L	14.24	9.97-18.95
	MYO	ng/mL	307.24	215.07-408.63
Auto Lumo A2000/A2000 plus	BNP	pg/mL	248.62	174.03-330.66
	CK-MB	ng/mL	101.04	70.73-134.39
	CRP	mg/L	4.30	3.01-5.72
	cTnl	ng/mL	1.66	1.16-2.20
	MYO	ng/mL	202.47	141.73-269.28
	NT-proBNP	pg/mL	514.80	360.36-684.68
Beckman Access 2	BNP	pg/mL	634.80	444.36-844.28
	CK-MB	ng/mL	71.15	49.81-94.63
	Digoxin	ng/mL	1.85	1.30-2.47
	MYO	ng/mL	180.72	126.51-240.36
Beckman Coulter AU680	CK	U/L	227.05	158.94-301.98
	CK-MB	U/L	33.79	23.65-44.94
i-CHROMATM Reader	CRP	mg/L	3.27	2.29-4.35
Siemens ADVIA Centaur XP	Digoxin	ng/mL	2.03	1.42-2.70
Roche Cobas C501	CRP	mg/L	3.96	2.77-5.15
	HCY	μmol/L	16.52	11.56-21.48
Roche Cobas E601	CK-MB	ng/mL	49.63	34.74-66.01
	cTnT	pg/mL	905.15	633.61-1203.85
	MYO	ng/mL	222.49	155.74-295.91
	NT-proBNP	pg/mL	471.80	330.26-627.50
Toshiba TBA-120FR	CK	U/L	241.60	169.12-321.33
	CK-MB	U/L	33.05	23.14-43.96
	CRP	mg/L	2.14	1.50-2.85
	HCY	μmol/L	14.67	10.27-19.51
	H-FABP	ng/mL	15.41	10.78-20.49

Remark: The range produced herein should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.