

## Cardiac Markers Control– Level 3

Assignment of values ( **LOT** 2022010601 )

Instrument	Analyte	Unit	Cardiac Markers Control-Level 3	
			Target	Range
Abbott I2000sr	BNP	pg/mL	742.63	519.84-965.42
	CK-MB	ng/mL	113.99	79.80-148.19
	cTnl	ng/mL	19.11	13.37-24.84
	HCY	µmol/L	34.32	24.02-44.61
	MYO	ng/mL	692.60	484.82-900.38
Auto Lumo A2000/A2000 plus	BNP	pg/mL	510.42	357.30-663.55
	CK-MB	ng/mL	193.91	135.74-252.09
	CRP	mg/L	9.56	6.70-12.43
	cTnl	ng/mL	15.90	11.13-20.67
	MYO	ng/mL	525.42	367.79-683.04
	NT-proBNP	pg/mL	7788	5451-10124
Beckman Access 2	BNP	pg/mL	1446	1013-1880
	CK-MB	ng/mL	138.11	96.68-179.54
	Digoxin	ng/mL	2.82	1.97-3.66
	MYO	ng/mL	470.70	329.49-611.91
Beckman Coulter AU680	CK	U/L	521.20	364.84-677.56
	CK-MB	U/L	63.26	44.28-82.24
i-CHROMATM Reader	CRP	mg/L	8.72	5.78-11.66
Siemens ADVIA Centaur XP	Digoxin	ng/mL	2.77	1.94-3.60
Roche Cobas C501	CRP	mg/L	9.12	6.38-11.86
	HCY	µmol/L	36.89	25.82-47.00
Roche Cobas E601	CK-MB	ng/mL	95.58	66.91-124.26
	cTnT	pg/mL	5412	3788-7035
	MYO	ng/mL	508.18	355.72-660.63
	NT-proBNP	pg/mL	7445	5212-9679
Toshiba TBA-120FR	CK	U/L	556.00	389.20-722.80
	CK-MB	U/L	62.01	43.40-80.61
	CRP	mg/L	6.52	4.56-8.47
	HCY	µmol/L	34.08	23.86-44.31
	H-FABP	ng/mL	30.92	21.64-40.20

Remark: The range produced herein should only be considered as a reference and it is recommended that each laboratory establish its own mean and acceptable ranges.