



TORCH IgG CONTROL——level4

Assignment of values (**LOT** 2022031801-A (06G24))

Instrument	Analyte	Unit	Level 4	
			Target	Range
AutoLumo A2000/A2000Plus	CMV-IgG	AU/mL	215.46	129.28~301.65
	RV-IgG	IU/mL	64.21	38.53~89.9
	TOX-IgG	IU/mL	12.93	7.76~18.1
	HSV-1IgG	AU/mL	216.66	129.99~303.32
	HSV-2IgG	AU/mL	261.02	156.61~365.43

Mark: The above quality control range is only for the reference of this batch of quality control products. The quality control range is batch specific. Each laboratory should determine the target value and range of quality control products according to the testing system and allowable limit used by itself, and establish the quality control range suitable for the reagents and instruments in the laboratory.